

<b>Measuring Range (1)</b> - Interval, determined by calibration, between the highest and the lowest content, where the lowest possible limit of the working range is the of quantification of the analytical method.	
<b>Procedure</b> - Statistic evaluation of method performance.	<b>Result</b> 0.00 – 1000 NTU

<b>Detection Limit (LOD) (2)</b> - The constituent concentration that, when processes through the complete method, produces a signal with 99% probability that it is different from the blank in reagent water that produces a signal above the mean of blank analyses.	
<b>Procedure</b> - 3 Standard deviation of 10 replicates by 1 NTU Certified Reference Material.	<b>Result</b> 0.15 NTU

<b>Quantification Limit (LOQ) (2)</b> - The constituent concentration that, when processes through the complete method, produces a signal sufficient greater than the blank that it can be detected within specified level by good laboratories during routine operating condition	
<b>Procedure</b> - 10 Standard deviation of 10 replicates by 1 NTU Certified Reference Material.	<b>Result</b> 0.50 NTU

<b>Uncertainty (of measurement) (3)</b> - Parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the measurand	
<b>Procedure (4)</b> – In compliance with QUAM-2012.P1	<b>Result</b> ± 55% (at 1 NTU) ± 20% (at 4 NTU) ± 10% (at 15 NTU) ± 6% (at 100 NTU) ± 6% (at 750 NTU)

CERTIFIED REFERENCE MATERIAL USED FOR UNCERTAINTY ESTIMATION	
VALUE NTU	1.00 – 10.00 – 100 - 750
MANUFACTURER	ISO 17034 REFERENCE MATERIAL PRODUCER

**Reference Document**

- (1): ISO 8466-1
- (2): Standard Methods for the Examination of Water and Waste water, 1010/1020
- (3): JCM 100 - Evaluation of measurement data — Guide to the expression of uncertainty in measurement
- (4): QUAM-2012.P1: Quantifying Uncertainty in Analytical Measurement